

1.1 Controlling the Module

Communication with the module is by standard VME bus protocols. All registers and memory locations are defined to be 4-byte entities. The VME slave module has three distinct address ranges.

A24 – The base address of this range is set by a 12-element DIP switch on the board. It occupies 4 Kbytes of VME address space, organized in 1 K 32-bit words. Relative to the base address, this space is divided as follows:

000-0FF – Register space to control and monitor the module (64 long words)

100-1FF – ADC processing registers (64 long words)

200-2FF – HITSUM processing registers (64 long words)

300-3FF – SCALER registers (64 long words)

400-4FF – SYSTEM TEST registers (64 long words)

500-FFF – Reserved (704 long words)

In addition to registers that are directly mapped to a VME address (Primary Address), the module supports Secondary Addressing in the A24 address space. These registers are accessed through an address mapping register (Secondary Address Register). Each secondary address is associated with a primary address. A Primary Address may have up to 64 K secondary addresses associated with it. A VME cycle loads the mapping register with data which is the internal (secondary) address of the target register. A VME cycle with the associated primary address accesses (read/write) the chosen internal register. Important registers are assigned primary addresses, allowing them to be directly accessible in a single VME cycle. Setup tables are assigned secondary addresses. This allows for a large internal address space, while maintaining a small VME footprint.

A32 - The base address of this range is programmed into register ADR32. It occupies 8 Mbytes of VME address space, organized in 2 M 32-bit words. A read of any address in this range will yield the next FADC data word from the module. Even though the module is logically a FIFO, the expanded address range allows the VME master to increment the address during block transfers. This address range can participate in single cycle, 32-bit block, and 64-bit block reads. The only valid write to this address range is the data value 0x80000000 which re-enables the module to generate interrupts (after one has occurred). The address range must be enabled by setting ADR32[0] = 1.

A32 - The lower and upper limits of this address range are programmed into register ADR_MB. This common address range for a set of FADC modules in the crate is used to implement the Multiblock protocol. By means of token passing FADC data may be read out from multiple FADC modules using a single logical block read. The board possessing the token will respond to a read cycle in this address range with the next FADC data word from that module. The token is passed along a private daisy chain line to the next module when it has transferred all data from a programmed number of events (register BLOCK SIZE). The address range must be enabled: set ADR_MB[0] = 1.

1.3 Module Registers

VERSION – board/firmware revision (0x0)

[7...0] – (R) – firmware revision

[15...8] – (R) – board revision

[31...16] – (R) – board type (“FADC”)

CSR – Control/Status (0x4)

0 – (R) – Event Accepted

1 – (R) – Block of Events Accepted

2 – (R) – Block of Events ready for readout

3 – (R) – BERR Status (1 = BERR asserted)

4 – (R) – Token Status (1 = module has token)

[5...10] – (reserved)

11 – (R) – Data FIFO Empty Flag Asserted

12 – (R) – Data FIFO Almost Empty Flag Asserted

13 – (R) – Data FIFO Half Full Flag Asserted

14 – (R) – Data FIFO Almost Full Flag Asserted

15 – (R) – Data FIFO Full Flag Asserted

- [16...19] – (reserved)
- 20 – (W) – Pulse Soft **Trigger 2** (if CTRL[7] = 1 and CTRL[6..4] = 5)
 (delayed **Trigger 1** follows; delay in TRIG21_DELAY register)
 (R) – Trigger 2 -> Trigger 1 sequence active
- 21 – (W) – Pulse Clear Module – soft reset + clear data pipelines
 (R) – Clear Module process active
- 22 – (W) – ENABLE SCALERS INTO DATA STREAM with FORCED
 BLOCK TRAILER INSERTION (write ‘1’ to bits 22, 23)
- 23 – (W) – FORCE BLOCK TRAILER INSERTION – will be successful only
 if there are NO triggers waiting to be processed
- 24 – (R) – Last FORCE BLOCK TRAILER INSERTION Successful
- 25 – (R) – Last FORCE BLOCK TRAILER INSERTION Failed
- 26 – (R) – Local Bus Time Out – target AK or DK timed out (5 us);
- 27 – (R/W) – Local Bus Error – target protocol violation;
 (write ‘1’ clears latched bits 26, 27)
- 28 – (W) – Pulse Soft **Sync Reset** (if CTRL[11] = 1 and CTRL[10..8] = 6)
- 29 – (W) – Pulse Soft **Trigger 1** (if CTRL[7] = 1 and CTRL[6..4] = 6)
- 30 – (W) – Pulse Soft Reset – initialize counters, state machines, memory
- 31 – (W) – Pulse Hard Reset – initialize module to power-up state

CTRL1 – Control 1 (0x8)

- [1...0] – (R/W) – Sampling Clock Source Select
 - 0 = Internal Clock
 - 1 = Front Panel connector
 - 2 = P0 connector (VXS)
 - 3 = P0 connector (VXS)
- 2 – (not used)
- 3 – (R/W) – Enable Internal Clock

[6...4] – (R/W) – Trigger Source Select

- 0 = Front Panel Connector (**Trigger 1**)
- 1 = Front Panel Connector (**Trigger 1**; synchronized)
- 2 = P0 Connector (VXS) (**Trigger1, Trigger 2**)
- 3 = P0 Connector (VXS) (**Trigger1, Trigger 2**; synchronized)
- 4 – (not used)
- 5 – Software Generated (**Trigger 2** + delayed **Trigger 1**)
- 6 = Software Generated (**Trigger 1**)
- 7 = Module Internal Logic

7 – (R/W) – Enable Soft Trigger

[10...8] – (R/W) – Sync Reset Source Select

- 0 = Front Panel Connector
- 1 = Front Panel Connector (synchronized)
- 2 = P0 Connector (VXS)
- 3 = P0 Connector (VXS) (synchronized)
- 4 – (not used)
- 5 – (not used)
- 6 = Software Generated
- 7 = no source

11 – (R/W) – Enable Soft Sync Reset

12 – (R/W) – Select Live Internal Trigger to Output

13 – (R/W) – Enable Front Panel Trigger Output

14 – (R/W) – Enable P0 (VXS) Trigger Output

[15...17] – (reserved)

18 – (R/W) – Enable Event Level Interrupt

19 – (reserved)

20 – (R/W) – Enable BERR response

21 – (R/W) – Enable Multiblock protocol

22 – (R/W) – FIRST board in Multiblock system

23 – (R/W) – LAST board in Multiblock system

24 – (reserved)

25 – (R/W) – Enable Debug Mode

[26...27] – (reserved)

28 – (R/W) – Multiblock Token passed on P0

29 – (R/W) – Multiblock Token passed on P2

30 – (reserved)

31 – (R/W) – System Test Mode (0 = normal, 1 = test mode enabled)

CTRL2 – Control 2 (0xC)

0 – (R/W) – GO (allow data transfer from external FIFOs to input FIFOs)

1 – (R/W) – Enable **Trigger (1 & 2)** to Module (source = CTRL1[6...4])

2 – (R/W) – Enable **Sync Reset** to Module (source = CTRL1[10...8])

3 – (R/W) – Enable Internal Trigger Logic

4 – (R/W) – Enable Streaming mode (NO event build)

[5...7] – (reserved)

8 – (R/W) – Enable Test Event Generation (for debug)

[9...15] – (reserved)

Bits 16 – 31 are functional only in Debug Mode (CTRL1[25] = 1)

16 – (reserved)

17 – (R/W) – Transfer data: build FIFO → output FIFO

[18...31] – (reserved)

BLOCK SIZE (0x10)

[15...0] – (R/W) – number of events in a BLOCK.

Stored Event Count \geq BLOCK SIZE → CSR[3] = 1.

[31...16] – (reserved)

INTERRUPT (0x14)

[7...0] – (R/W) – Interrupt ID (vector)

[10...8] – (R/W) – Interrupt Level [2..0]. Valid values = 1,..,7.

11 - 15 – (reserved)

[20...16] – (R) – Geographic Address (slot number) in VME64x chassis.

21 – 22 – (reserved)

23 – (R) – Parity Error in Geographic Address.

24 – 31 – (reserved)

ADR32 – Address for data access (0x18)

0 – (R/W) – Enable 32-bit address decoding

1 – 6 – (reserved – read as 0)

[15...7] – (R/W) – Base Address for 32-bit addressing mode (8 Mbyte total)

ADR MB – Multiblock Address for data access (0x1C)

0 – (R/W) – Enable Multiblock address decoding

1 – 6 – (reserved – read as 0)

[15...7] – (R/W) – Lower Limit address (ADR_MIN) for Multiblock access

16 – 22 – (reserved – read as 0)

[31...23] – (R/W) – Upper Limit address (ADR_MAX) for Multiblock access

The board that has the TOKEN will respond with data when the VME address satisfies the following condition:

$$\text{ADR_MIN} \leq \text{Address} < \text{ADR_MAX}.$$

SEC_ADR – Secondary Address (0x20)

[15...0] – (R/W) – Secondary Address for 24-bit addressing mode

16 – (R/W) – Enable auto-increment mode (secondary address increments by 1 after each access of the associated primary address)

DELAY – Trigger/Sync Reset Delay (0x24)

[21...16] – (R/W) – Sync reset delay

[5...0] – (R/W) – Trigger delay

INTERNAL TRIGGER CONTROL (0x28)

[23...16] – (R/W) – trigger width (4 ns per count)

[7...0] – (R/W) – trigger hold off delay (4 ns per count)

RESET CONTROL (0x2C)

0 – (W) – Hard reset – Control FPGA

1 – (W) – Hard reset – ADC processing FPGA

[2...3] – (reserved)

4 – (W) – Soft reset – Control FPGA

5 – (W) – Soft reset – ADC processing FPGA

[6...7] – (reserved)

8 – (W) – Reset – ADC data FIFO

[9...10] – (reserved)

10 – (W) – Reset – HITSUM FIFO

11 – (W) – Reset – DAC (all channels)

12 – (W) – Reset – EXTERNAL RAM Read & Write Address Pointers

[13...15] – (reserved)

16 – (W) – Take Token – return token to 1st board of multiboard set

[17...31] – (reserved)

TRIGGER COUNT (0x30)

[31...0] – (R) – total trigger count

31 – (W) – reset count

EVENT COUNT (0x34)

[23...0] – (R) – number of events on board (non-zero → CSR[0] = 1).

[31...24] – (reserved)

BLOCK COUNT – (0x38)

[31...20] – reserved

[19...0] – (R) - number of event BLOCKS on board (non-zero → CSR[2] = 1).

BLOCK FIFO COUNT – (0x3C)

[31...6] – reserved

[5...0] – (R) - number of entries in BLOCK WORD COUNT FIFO

BLOCK WORD COUNT FIFO – (64 deep FIFO) (0x40)

[31...25] – reserved (read as ‘0’)

24 – (R) – count not valid (word count FIFO empty)

[23...20] – reserved (read as ‘0’)

[19...0] – (R) - number of words in next event BLOCK

INTERNAL TRIGGER COUNT (0x44)

[31...0] – (R) – internal live trigger count

31 – (W) – reset count

EXTERNAL RAM WORD COUNT (0x48)

[31...22] – reserved (read as ‘0’)

21 – (R) – RAM empty

20 – (R) – RAM full (1,048,576 eight byte words)

[19...0] – (R) – data word count (eight byte words)

DATA FLOW STATUS (0x4C) (for debug)

DAC 1_2 – DAC channels 1,2 (0x50)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16] – (R/W) – DAC value channel 1

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 2

DAC 3_4 – DAC channels 3,4 (0x54)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16] – (R/W) – DAC value channel 3

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 4

DAC 5_6 – DAC channels 5,6 (0x58)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16]– (R/W) – DAC value channel 5

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 6

DAC 7_8 – DAC channels 7,8 (0x5C)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16]– (R/W) – DAC value channel 7

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 8

DAC 9_10 – DAC channels 9,10 (0x60)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16]– (R/W) – DAC value channel 9

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 10

DAC 11_12 – DAC channels 11,12 (0x64)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16]– (R/W) – DAC value channel 11

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 12

DAC 13_14 – DAC channels 13,14 (0x68)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...18]– (R/W) – DAC value channel 13

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 14

DAC 15_16 – DAC channels 15,16 (0x6C)

31 – (reserved)

[30...28] – (reserved – read as 0)

[27...16] – (R/W) – DAC value channel 15

15 – (reserved)

[14...12] – (reserved – read as 0)

[11...0] – (R/W) – DAC value channel 16

STATUS 1 – Input Buffer Status (0x70)

31 – (R) – data buffer ready for input

30 – (R) – data buffer input paused

29 – (R) – reserved (read as ‘0’)

28 – (R) – data buffer empty

27 – (R) – data buffer full

[26...16] – (R) – data buffer word count

[15...0] – (reserved)

STATUS 2 – Build Buffer Status (0x74)

[31...29] – reserved (read as ‘0’)

28 – (R) – data buffer ‘A’ empty

27 – (R) – data buffer ‘A’ full

[26...16] – (R) – data buffer ‘A’ word count

[15...13] – reserved (read as ‘0’)

12 – (R) – data buffer ‘B’ empty

11 – (R) – data buffer ‘B’ full

[10...0] – (R) – data buffer ‘B’ word count

STATUS 3 – Output Buffer Status (0x78)

[31...30] – reserved (read as ‘0’)

29 – (R) – data buffer ‘A’ empty

- 28 – (R) – data buffer ‘A’ full
- [27...16] – (R) – data buffer ‘A’ word count
- [15...14] – reserved (read as ‘0’)
- 13 – (R) – data buffer ‘B’ empty
- 12 – (R) – data buffer ‘B’ full
- [11...0] – (R) – data buffer ‘B’ word count

STATUS 4 – (spare) (0x7C)

- [31...0] – reserved

AUXILIARY 1 – (spare) (0x80)

- [31...0] – reserved

AUXILIARY 2 – (spare) (0x84)

- [31...0] – reserved

TRIG21 DELAY (0x88)

- [31...12] – reserved

- [11...0] – (R/W) – Delay from soft TRIG2 to generated TRIG1 (4 ns/count)

RAM Address Register (0x8C) – The RAM is organized as two 36-bit words with a common address. Auxiliary VME access (R/W) to the RAM is provided through a pair of 32 bit data registers (RAM 1, RAM 2). Note that bits 35 – 32 of each RAM word are not accessible through VME. During data flow operations, these bits carry event marker tags (header, trailer).

- 31 – increment address after access (R/W) of RAM 1 Data Register
- 30 – increment address after access (R/W) of RAM 2 Data Register

[29...21] – reserved (read as 0)

[19...0] – RAM address

RAM 1 Data Register (0x90)

[31...0] – RAM data word bits 67 – 36 (32 bits)

RAM 2 Data Register (0x94)

[31...0] – RAM data word bits 31 – 0 (32 bits)

(PROM Registers 1 and 2 are used for FPGA configuration over VME.)

PROM Register 1 (0x98)

31 – READY – (R) – configuration state machine is available to accept command
(i.e. no configuration process is currently executing).

[30...8] – reserved (read as 0)

[7...0] – configuration OPCODE

PROM Register 2 (0x9C)

[31...0] – PROM ID – (R) response to specific OPCODE write to PROM reg 1.

BERR Module Count (0xA0)

[31...0] – BERR count (driven by module to terminate data transmission)

BERR Total Count (0xA4)

[31...0] – BERR count (as detected on bus)

Auxiliary Scaler 1 (0xA8)

[31...0] – Total word count from ADC Processing FPGA

Auxiliary Scaler 2 (0xAC)

[31...0] – (reserved)

Auxiliary Scaler 3 (0xB0)

[31...0] – Event header word count from ADC Processing FPGA

TRIGGER 2 SCALER (0xB4)

[31...0] – (R) – **Trigger 2** count

31 – (W) – write ‘1’ to reset count

Auxiliary Scaler 5 (0xB8)

[31...0] – Event trailer word count from ADC Processing FPGA

SYNC RESET SCALER (0xBC)

[31...0] – (R) – **Sync Reset** count

31 – (W) – write ‘1’ to reset count

Module Busy Level (0xC0)

[31] – Force module busy

[30...20] – reserved

[19...0] – Busy level (eight byte words)

(External RAM word count > Busy level → module busy = 1)

Generate Event Header Word (0xC4) (for debug)

[31...0] – (W) – Event Header Word

Generate Event Data Word (0xC8) (for debug)

[31...0] – (W) – Event Data Word

Generate Event Trailer Word (0xCC) (for debug)

[31...0] – (W) – Event Trailer Word

MGT STATUS (0xD0)

0 – (R) – lane 1 up (GTX1)

1 – (R) – lane 2 up (GTX1)

2 – (R) – channel up (GTX1)

3 – (R) – hard error (GTX1)

4 – (R) – soft error (GTX1)

5 – (R) – lane 1 up (GTX2)

6 – (R) – lane 2 up (GTX2)

7 – (R) – channel up (GTX2)

8 – (R) – hard error (GTX2)

9 – (R) – soft error (GTX2)

10 – (R) – SUM DATA VALID

11 – (R) – MGT RESET ASSERTED

[31...12] – (R) - Reserved

MGT CONTROL (0xD4)

0 – **RELEASE MGT RESET** (0 = reset MGT, 1 = release reset)

1 – Data Type to CTP (0 = counting sequence, 1 = front-end data)

2 – Enable Data Alignment on Sync Reset occurrence

[31...3] – Reserved

RESERVED (2 registers) (0xD8 – 0xDC)

SCALER CONTROL (0xE0) – See SCALERS (0x300 – 0x340)

0 – (R/W) – Enable all scalers to count (1 = enable, 0 = disable)

1 – (W) – Latch all scalers. Write ‘1’ to simultaneously transfer all 17 scaler counts to registers for readout.

2 – (W) – Reset all scalers. Write ‘1’ to simultaneously reset all 17 scaler counts to zero.

[3 – 31] – (reserved)

BOARD SERIAL NUMBER 0 (0xE4)

[31...24] – (R) – board serial number byte 0

[23...16] – (R) – board serial number byte 1

[15...8] – (R) – board serial number byte 2

[7...0] – (R) – board serial number byte 3

BOARD SERIAL NUMBER 1 (0xE8)

[31...24] – (R) – board serial number byte 4

[23...16] – (R) – board serial number byte 5

[15...8] – (R) – board serial number byte 6

[7...0] – (R) – board serial number byte 7

BOARD SERIAL NUMBER 2 (0xEC)

[31...24] – (R) – board serial number byte 8

[23...16] – (R) – board serial number byte 9

[15...8] – (R) – board serial number byte 10

[7...0] – (R) – board serial number byte 11

SCALER INSERTION INTERVAL (0xF0) - Data from the SCALERS defined below (0x300 – 0x340) may be inserted into the readout data stream at regular event count intervals. The interval is specified in multiples of the event BLOCK SIZE. When the interval is ZERO (the default condition), there is NO insertion of scaler data into the data stream. When programmed for a non-zero interval, the current scaler values are appended to the last event of the appropriate BLOCK of events. The current Trigger 1 count is also inserted as the 18th scaler. Note that the scalers are **NOT** reset after their values are captured.

Example: Interval = 10 means that every 10th block of events will have the integrated scaler data appended to it.

(See the document FADC V2 Data Format for information on identifying scaler data words in an event.)

The scalers may ALSO be inserted into the data stream when a FORCE BLOCK TRAILER is done by the user. A simultaneous write of ‘1’ to bit 22 and bit 23 of the CSR (0x4) accomplishes this. The scaler values are those at the time of the last trigger’s occurrence.

[15...0] - (R/W) – N (in BLOCKS of events); every Nth block of events has integrated scaler data appended to the last event in the block.

[31...16] – (reserved)

SUM THRESHOLD (History buffer) (0xF4)

31 – (R) – sum data READY for readout if value = ‘1’

[30...16] – (R/W) – reserved (read as ‘0’)

[15...0] – (R/W) – sum threshold value for data capture

SUM DATA (History buffer) (0xF8)

31 – (W) – writing a ‘1’ ARMs History buffer for data capture

[30...16] – (R) – reserved (read as ‘0’)

[15...0] – (R) – sum data sample

SPARE (1 register) (0xFC)

SCALER Registers (0x300 – 0x340) (R)

SCALER[0] – (0x300) - input channel 0 count

SCALER[1] – (0x304) - input channel 1 count

SCALER[2] – (0x308) - input channel 2 count

SCALER[3] – (0x30C) - input channel 3 count

SCALER[4] – (0x310) - input channel 4 count

SCALER[5] – (0x314) - input channel 5 count

SCALER[6] – (0x318) - input channel 6 count

SCALER[7] – (0x31C) - input channel 7 count

SCALER[8] – (0x320) - input channel 8 count

SCALER[9] – (0x324) - input channel 9 count

SCALER[10] – (0x328) - input channel 10 count

SCALER[11] – (0x32C) - input channel 11 count

SCALER[12] – (0x330) - input channel 12 count

SCALER[13] – (0x334) - input channel 13 count

SCALER[14] – (0x338) - input channel 14 count

SCALER[15] – (0x33C) - input channel 15 count

TIME COUNT – (0x340) - timer (each count represents 2048 ns)

System Test Registers (0x400 – 0x410)

TEST BIT REGISTER (0x400)

- 0 – (R/W) – trigger_out_p0 (1 = asserted, 0 = not asserted)
- 1 – (R/W) – busy_out_p0 (1 = asserted, 0 = not asserted)
- 2 – (R/W) – sdlink_out_p0 (1 = asserted, 0 = not asserted)
- 3 – (R/W) – token_out_p0 (1 = asserted, 0 = not asserted)
- [4 – 7] – (R/W) – spare out test bits
- 8 – (R) – status_b_in_p0 state (1 = asserted, 0 = not asserted)
- 9 – (R) – token_in_p0 state (1 = asserted, 0 = not asserted)
- [10 - 14] – (R) – reserved (read as ‘0’)
- 15 – (R) – clock_250 counter status (1 = counting, 0 = not counting)
- [16 - 31] – (R) – reserved (read as ‘0’)

CLOCK_250 COUNT REGISTER (0x404)

- 0 – (W) – Write ‘0’ resets the counter. Write ‘1’ initiates 20us counting interval.
- [31 - 0] – (R) – CLK_250 counter value. (Should be 5000 after count interval.)

SYNC_IN_P0 COUNT REGISTER (0x408)

- 0 – (W) – Write ‘0’ resets the counter.
- [31 - 0] – (R) – SYNC_IN_P0 counter value.

TRIG1_IN_P0 COUNT REGISTER (0x40C)

- 0 – (W) – Write ‘0’ resets the counter.
- [31 - 0] – (R) – TRIG1_IN_P0 counter value.

TRIG2_IN_P0 COUNT REGISTER (0x410)

0 – (W) – Write ‘0’ resets the counter.

[31 - 0] – (R) – TRIG2_IN_P0 counter value.

ADC PROCESSING FPGA ADDRESS MAP:

Control Bus Memory Map for FADC FPGA

Name [VME ADDRESS]	Width (Bits)	Quantity	Access	Primary Address (Secondary Address)	Function
STATUS0 [0x100]	16	1	R	0x0000 (---)	Bits 14 to 0: Code Version Bit 15: 1= Command can be sent to AD9230
STATUS1 [0x104]	16	1	R	0x0001 (---)	TRIGGER NUMBER BIT 15 to 0
STATUS2 [0x108]	16	1	R	0x0002 (---)	Tbd. Read 0
CONFIG 1 [0x10C]	16	1	R/W	0x0003 (---)	Bit 0-2 (process mode): 000 → Select option1 001 → Select option2 010 → Select option3 011 → Select option4 111 → Run option1 then option4 for each trigger Bit 3: 1:Run Bit 6-5 : Number of Pulses in Mode 1 and 2 Bit 7: Test Mode (play Back).
CONFIG 2 [0x110]			R/W	0x0004 (---)	When 1 ADC values = 0 Bit 0 → ADC 0 Bit 1 → ADC 1 Bit 2 → ADC 2 Bit 3 → ADC 3 Bit 4 → ADC 4 Bit 5 → ADC 5 Bit 6 → ADC 6 Bit 7 → ADC 7 Bit 8 → ADC 8 Bit 9 → ADC 9 Bit 10 → ADC 10

					Bit 11→ ADC 11 Bit 12→ ADC 12 Bit 13→ ADC 13 Bit 14→ ADC 14 Bit 15→ ADC 15
CONFIG 4 [0x114]	16	1		0x0005 (---)	7 => rising edge write to AD9230 ADC 6 => 1 write to all ADC. Bits 3..0 are don't care 5 => 0 write to AD9230 1 read from AD9230 3..0 => Select ADC to write to
CONFIG 5 [0x118]	16	1		0x0006 (---)	15..8 => Registers inside AD9230 7..0 => Data to write to register.
PTW [0x11C]	9	1	R/W	0x0007 (---)	Number of ADC sample to include in trigger window. PTW = Trigger Window (ns) * 250 MHz. Minimum is 6. Always report Even Number. For odd PTW number, discard the last sample reported.
PL [0x120]	11	1		0x0008 (---)	Number of sample back from trigger point. PL = Trigger Window(ns) * 250MHz
NSB [0x124]	12	1		0x0009 (---)	Number of sample before trigger point to include in data processing. This include the trigger Point. Minimum is 2 in all mode.
NSA [0x128]	13	1		0x000A (---)	Number of sample after trigger point to include in data processing. Minimum is (6 in mode 2)and (3 in mode 0 and 1). Number of

					sample report is 1 more for odd and 2 more for even NSA number.
TET [0x12C – 0x148] (2 channels per word) (see Note 1 below)	12	16		0x000B - 0x001A (---)	Trigger Energy Threshold.
PTW DAT BUF LAST ADR [0x14C]	12	1		0x001B (---)	Last Address of the Secondary Buffer. See calculation below
PTW MAX BUF [0x150]	8	1		0x001C (---)	The maximum number of unprocessed PTW blocks that can be stored in Secondary Buffer. See Calculation below.
Test Wave Form [0x154]	16	1		0x001D (---)	Write to PPG. Read should immediately follow write.
ADC0 Pedestal Subtract [0x158]	16	1	R/W	0x001E	Subtract from ADC0 Count before Summing
ADC1 Pedestal Subtract [0x15C]	16	1	R/W	0x001F	Subtract from ADC1 Count before Summing
ADC2 Pedestal Subtract [0x160]	16	1	R/W	0x0020	Subtract from ADC2 Count before Summing
ADC3 Pedestal Subtract [0x164]	16	1	R/W	0x0021	Subtract from ADC3 Count before Summing
ADC4 Pedestal Subtract [0x168]	16	1	R/W	0x0022	Subtract from ADC4 Count before Summing
ADC5 Pedestal Subtract [0x16C]	16	1	R/W	0x0023	Subtract from ADC5 Count before Summing
ADC6 Pedestal Subtract [0x170]	16	1	R/W	0x0024	Subtract from ADC6 Count before Summing
ADC7 Pedestal Subtract [0x174]	16	1	R/W	0x0025	Subtract from ADC7 Count before Summing

ADC8 Pedestal Subtract [0x178]	16	1	R/W	0x0026	Subtract from ADC8 Count before Summing
ADC9 Pedestal Subtract [0x17C]	16	1	R/W	0x0027	Subtract from ADC9 Count before Summing
ADC10 Pedestal Subtract [0x180]	16	1	R/W	0x0028	Subtract from ADC10 Count before Summing
ADC11 Pedestal Subtract [0x184]	16	1	R/W	0x0029	Subtract from ADC11 Count before Summing
ADC12 Pedestal Subtract [0x188]	16	1	R/W	0x002A	Subtract from ADC12 Count before Summing
ADC13 Pedestal Subtract [0x18C]	16	1	R/W	0x002B	Subtract from ADC13 Count before Summing
ADC14 Pedestal Subtract [0x190]	16	1	R/W	0x002C	Subtract from ADC14 Count before Summing
ADC15 Pedestal Subtract [0x194]	16	1	R/W	0x002D	Subtract from ADC15 Count before Summing

$$PTW \text{ MAX BUF} = \text{INT}(2016 / (PTW + 8) * 250000000)$$

Where:

2016 → Number of address of Secondary Buffer

PTW → Trigger Window width in nano-second

$$PTW \text{ DAT BUF LAST ADR} = PTW \text{ MAX BUF} * (PTW + 6) - 1;$$

Where:

6 → 4 address for Time Stamp and 2 address for Trigger Number

NumberOfBytePerTrigger → PTW * 250 MHz.

NOTE 1: Trigger Energy Threshold (TET)

0x12C – Channel 1 & Channel 2

- [31...28] – not used
- [27...16] – channel 1 threshold
- [15...12] – not used
- [27...16] – channel 2 threshold

0x130 – Channel 3 & Channel 4

- [31...28] – not used
- [27...16] – channel 3 threshold
- [15...12] – not used
- [27...16] – channel 4 threshold

.....

0x148 – Channel 15 & Channel 16

- [31...28] – not used
- [27...16] – channel 15 threshold
- [15...12] – not used
- [27...16] – channel 16 threshold